

SRD1000 Superconductive Reference Devices

Reporting period: mid-2002 - December 2003

Presentation at CNRS-CRTBT, Grenoble, 11 December 2003

Wim Bosch, HDL, Leiden, The Netherlands

HDLinfo@xs4all.nl

www.xs4all.nl/~hdleiden/srd1000

SRD1000 Superconductive Reference Devices

Development

HDL

LION-KOL

NMi-VSL

UT

Evaluation



BNM-INM

PTB


NPL

CNRS / AL

SRD1000 introduction







SRD1000 introduction



#	material	T_c	W_c	uncert
		[mK]	[mK]	[%]
1	W	15	0.2	0.9
2	Be	21.98	0.08	0.5
3	Ir ₈₀ Rh ₂₀	32.32	0.99	0.4
4	Ir ₉₂ Rh ₀₈	66.1	0.94	0.2
5	Ir	99.34	0.55	0.1
6	AuAl ₂	160.89	0.73	0.2
7	AuIn ₂	207.80	1.56	0.2
8	Cd	516.9	14	0.2
9	Zn	843.7	5.9	0.1
10	Al	1166	4.0	<0.1

(values found for SRD004)

Residual field measurement and compensation



T_C of AuAl_2 versus I_{comp} of SRD004, as measured by PTB.


Residual field measurement and compensation

- ACS-10 adjustable DC current source
- Optional input on MIDS-10 preamp for compensation field in X,Y plane
- DCS-10 degauss tool for the Cryoperm shield




Calibration of prototypes at NMI-VSL

- run 10 (June 2002): SRD003 / 004 / 005
 - SRD003 / 004: new samples and detectors required
 - SRD005: sent to BNM-INM for evaluation
- run 11 (October 2002): SRD003 / 004 / 006 / 007
 - $T_{\text{base}} > 16 \text{ mK}$, no W calibration possible
 - occasionally the MCT blocks
 - temperature stabilisation problematic around 1 K
 - RhFe thermometer is not working properly
 - SRD003: AuAl_2 is not visible, repair detector before evaluation
 - SRD004: sent to PTB for evaluation
 - SRD006/007: AuIn_2 has long tail, new sample before evaluation




Calibration of prototypes at NMI-VSL

Berillium (run 11)




Calibration of prototypes at NMi-VSL

Ir92Rh08 (run 11)




Calibration of prototypes at NMi-VSL

Ir80Rh20 (run 11)




Calibration of prototypes at NMI-VSL

Iridium (run 11)




Calibration of prototypes at NMi-VSL

AuAl2 (run 11)




Calibration of prototypes at NMI-VSL

AuIn2 (run 11)




Calibration of prototypes at NMi-VSL

Cadmium (run 11)




Calibration of prototypes at NMi-VSL

Zinc (run 11)



Calibration of prototypes at NMI-VSL

Aluminium (run 11)





Improvements to samples after run 11



- AuAl₂
 - annealing, pretest samples before mounting in prototypes
 - new samples were mounted in SRD003 and SRD006
- AuIn₂
 - other batch produces better samples, pretest samples
 - new sample was mounted in SRD006
- Cd
 - etching the surface reduces W_c by factor 3
- Zn
 - etching gives no improvement
- Al
 - annealed wire replaces foil to reduce upturn effect bond wires
 - new samples were mounted in SRD003 and SRD006

Improvements to samples after run 11

AuAl₂





Improvements to samples after run 11

 AuIn_2 



Improvements to samples after run 11

Cd




Improvements to samples after run 11

Zn




Improvements to samples after run 11

Al



SRD005 / BNM



SRD006 / NPL

Evaluation of prototypes

#	calibration	samples modified	evaluation	status
SRD003	run 11	AuAl ₂ , Al	CNRS / AL	in progress
SRD004	run 11	-	PTB	report ready
SRD005	run 10	-	BNM-INM	report ready
SRD006	run 11	AuIn ₂ , AuAl ₂ , Al	NPL	in progress

2003

- Peruzzi, A., de Groot, M.J. , "Evaluation of Uncertainty in the Realization of the Provisional Low Temperature Scale from 10 mK to 1 K at Nmi" in *Proceedings of the International Conference on the Uncertainty of Measurement UNCERT 2003*);
- SRD partners, 'Report on the development of SRD technology', May 2003, project deliverable 2.1;
- HDL, 'Information on the installation of an SRD1000 device in a dilution refrigerator', June 2003;
- HDL, 'Procedure to degauss the SRD1000 Cryoperm shielding using the DCS-10', June 2003;
- HDL, 'Checking the residual magnetic field inside the SRD1000 shielding at low temperatures using the ACS10', June 2003;
- Peruzzi, A., Bosch, W.A., 'Procedure for the Evaluation of SRD1000 prototypes', July 2003;
- Presentation by Andrea Peruzzi at the Wroclaw Low Temperature Thermometry seminar, October 2003;
- Presentation by Andrea Peruzzi at Estec / ESA Noordwijk, December 2003.

2002

- Bosch, W.A., et al. , "First Prototypes of the Superconductive Reference Device SRD1000", in *Proceedings of the 8th Symposium on Temperature: Its Measurement and Control in Science and Industry*, Chicago, 2002;
- Bosch, W.A., et al., "SRD1000: a Superconductive Reference Device for thermometry below 1K", poster presented at the LT23 conference Japan, 2002, and *Physica B*, 329-333 (2003), pp. 1562-1563.

- Reports: evaluation, TIP, final
- Market developmet
- Resources for calibrations
- Pilot series (5 - 10 devices)
- SRD technology
 - Improve sample quality, additional batches (Ir / Ir_xRh_y)
 - Mo single crystal ($T_C = 900$ mK) to replace Zn / Al
 - Ti layer with $T_C = 300$ mK
 - Include thermometer for interpolation (CMN / PdFe / AuEr)